



ಕರ್ನಾಟಕ ಸರ್ಕಾರ

ರಾಷ್ಟ್ರೀಯ ಆರೋಗ್ಯ ಅಭಿಯಾನ

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ (೦)

ಜಿಲ್ಲಾ ಯೋಜನಾ ನಿರ್ವಹಣಾ ಘಟಕ, ಪಿಡಬ್ಲ್ಯುಡಿ ಬಿಲ್ಡಿಂಗ್, ನೆಹರೂ ಮೈದಾನ್ ರಸ್ತೆ ದ.ಕ. ಮಂಗಳೂರು

ಇಮೇಲ್: dpdodakshinakannada@gmail.com

ಫೋನ್ ನಂಬರ್: 0824-2424501

ನಂ. DPMU/DK/BPHU /15th FC/94D/2025-26

ದಿನಾಂಕ: 11/02/2026

Manual Tender Form

**TENDER FOR PHLEBOTOMY CHAIR,
THREE BUCKET MOP TROLLY SYSTEM
AND BIO-MEDICAL WASTE
SEGREGATION TROLLY FOR BPHUs OF
DAKSHINA KANNADA DISTRICT**



ಕರ್ನಾಟಕ ಸರ್ಕಾರ

ರಾಷ್ಟ್ರೀಯ ಆರೋಗ್ಯ ಅಭಿಯಾನ

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ (ರಿ)

ಜಿಲ್ಲಾ ಯೋಜನಾ ನಿರ್ವಹಣಾ ಘಟಕ, ಪಿಡಬ್ಲ್ಯುಡಿ ಬಿಲ್ಡಿಂಗ್, ನೆಹರೂ ಮೈದಾನ್ ರಸ್ತೆ ದ.ಕ. ಮಂಗಳೂರು

ಇಮೇಲ್: dpdodakshinakannada@gmail.com

ಫೋನ್ ನಂಬ್ರ: 0824-2424501

ನಂ. DPMU/DK/BPHU /15th FC/94D/2025-26

ದಿನಾಂಕ: 11/02/2026

Manual ಟೆಂಡರ್ ಪ್ರಕಟಣೆ

ದಕ್ಷಿಣ ಕನ್ನಡ ಜಿಲ್ಲೆಯ 15ನೇ ಹಣಕಾಸು ಯೋಜನೆಯಡಿ ಸುಳ್ಳುಬಂಟ್ವಾಳ, ಪುತ್ತೂರು ಮತ್ತು ಬೆಳ್ತಂಗಡಿ ತಾಲೂಕು ಆಸ್ಪತ್ರೆಯಲ್ಲಿ ಪ್ರಾರಂಭಗೊಳ್ಳಲಿರುವ Block Public Health Unit(BPHU)ಗಳಿಗೆ ಅಗತ್ಯವಿರುವ Phlebotomy chair, Three bucket mop trolley system ಮತ್ತು Bio-Medical waste segregation trolleyಗಳನ್ನು ಪೂರೈಸಲು ಅರ್ಹ ಸಂಸ್ಥೆಗಳಿಂದ Manual ಟೆಂಡರ್ ಆಹ್ವಾನಿಸಲಾಗಿದೆ.

ಅರ್ಹ ಸಂಸ್ಥೆಗಳು ಈ ಕಛೇರಿಯ ವೆಬ್‌ಸೈಟ್ www.dkhfw.in ತಂತ್ರಾಂಶದಲ್ಲಿ Manual ಟೆಂಡರ್ ನಮೂನೆಯನ್ನು ಪಡೆದು ಭರ್ತಿಗೊಳಿಸಿದ Manual ಟೆಂಡರ್ ನಮೂನೆಯನ್ನು Technical Documentನಲ್ಲಿ ತಿಳಿಸಲಾದ ಎಲ್ಲಾ ದಾಖಲಾತಿಗಳ ಪ್ರತಿಯೊಂದಿಗೆ The Member Secretary, DH&FWS ,Dakshina Kannada District (ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ, ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ, ಸೊಸೈಟಿ, ದ.ಕ ಮಂಗಳೂರು) ಇವರ ಹೆಸರಿನಲ್ಲಿ ಮೊಹರಾದ ಲಕೋಟೆಯಲ್ಲಿಟ್ಟು ದಿನಾಂಕ 17/02/2026ರ ಪೂರ್ವಾಹ್ನ 11.00 ಗಂಟೆಯೊಳಗಾಗಿ ಈ ಕಛೇರಿಗೆ ಸಲ್ಲಿಸತಕ್ಕದ್ದು.

ಸಹಿ/-

ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ, ಸೊಸೈಟಿ

ದ.ಕ ಮಂಗಳೂರು.

**Supply of Phlebotomy chair ,Three bucket mop trolley system & Bio-Medical
waste segregation trolley under BPHU Programme**

Technical Document

Sl. No	Particulars	Details to be Filled by Tenderer (Attached/not attached)
1	Name of the Firm, Firms full address Office Phone, Fax, and Mobile No and E-mail ID (In Firm's letter head)	
2	Attested Copy of Valid Firm Registration/Valid Trade License from concerned Authority	
3	Attested copy of GST certificate	
4	Attested copy of TAN or PAN Card	
5	Technical literature of the product quoted(brochures)	

ಟೆಂಡರ್‌ದಾರರ ಸಹಿ ಮತ್ತು ಮೊಹರು

ದಕ್ಷಿಣ ಕನ್ನಡ ಜಿಲ್ಲೆಯ 15ನೇ ಹಣಕಾಸು ಯೋಜನೆಯಡಿ ಸುಳ್ಳುಬಂಟ್ವಾಳ, ಪುತ್ತೂರು ಮತ್ತು ಬೆಳ್ತಂಗಡಿ ತಾಲೂಕು ಆಸ್ಪತ್ರೆಯಲ್ಲಿ ಪ್ರಾರಂಭಗೊಳ್ಳಲಿರುವ Block Public Health Unit(BPHU)ಗಳಿಗೆ ಅಗತ್ಯವಿರುವ Phlebotomy chair, Three bucket mop trolley system ಮತ್ತು Bio-Medical waste segregation trolleyಗಳನ್ನು ಪೂರೈಸಲು ನಿಬಂಧನೆಗಳು

- 1) ಅಂದಾಜು ವೆಚ್ಚ ಮೊತ್ತ ರೂ.1,30,000/- ಆಗಿರುತ್ತದೆ.
- 2) Manual ಟೆಂಡರ್ ಪರಿಶೀಲನೆಯ ಬಗ್ಗೆ ಮತ್ತು ಈ Manual ಟೆಂಡರ್ ಬಗ್ಗೆ ಇತರೆ ಯಾವುದೇ ದೂರುಗಳಿದ್ದಲ್ಲಿ ನದರೀ ಪ್ರಕ್ರಿಯೆ ನಡೆದ 1 ವಾರದೊಳಗೆ ಮಾತ್ರವೇ ಮೇಲ್ಮನವಿ ಪ್ರಾಧಿಕಾರಕ್ಕೆ ಮನವಿ ಸಲ್ಲಿಸಬೇಕು. ತದನಂತರ ಬಂದ ಯಾವುದೇ ಮನವಿಯನ್ನು ಪರಿಗಣಿಸಲಾಗುವುದಿಲ್ಲ.
- 3) ಯಶಸ್ವಿ ಸಂಸ್ಥೆಯು/ಇಲಾಖೆಯು ಒಪ್ಪಂದದಿಂದ ಹಿಂದೆ ಸರಿಯುವುದಾದರೆ 7 ದಿನಗಳ ಮುಂಚಿತವಾಗಿ ನೋಟೀಸ್‌ನ್ನು ಬರವಣಿಗೆಯಲ್ಲಿ ನೀಡತಕ್ಕದ್ದು.ಇಲ್ಲದಿದ್ದಲ್ಲಿ ಅಗತ್ಯ ಕ್ರಮ ಕೈಗೊಳ್ಳಲಾಗುವುದು.
- 4) ಸಂಸ್ಥೆಯವರು ಸೂಕ್ತವಾಗಿ Manual ಟೆಂಡರ್ ಸಲ್ಲಿಕೆ ಮಾಡಬೇಕು.ಯಾವುದೇ ತಪ್ಪುಗಳು ಸಂಭವಿಸಿದ್ದಲ್ಲಿ ಇಲಾಖೆಯು ಜವಾಬ್ದಾರಿ ಆಗಿರುವುದಿಲ್ಲ.
- 5) ಯಾವುದೇ ಸೂಚನೆಗಳನ್ನು ನೀಡದೇ Manual ಟೆಂಡರ್ ಅನ್ನು ರದ್ದುಪಡಿಸುವ/ಅಂಗೀಕರಿಸುವ/ತಿರಸ್ಕರಿಸುವ ಹಕ್ಕು ಇಲಾಖೆಗೆ ಇರುತ್ತದೆ.
- 6) Manual ಟೆಂಡರ್‌ನಲ್ಲಿ ನಮೂದಿಸುವ ದರವನ್ನು ಮಾತ್ರ ಪರಿಗಣಿಸಲಾಗುವುದು. ಆದುದರಿಂದ ಸಂಸ್ಥೆಯವರು ದರಗಳನ್ನು ಸೂಚಿಸುವಾಗ ಸಾಗಾಟದ ವೆಚ್ಚ, ತೆರಿಗೆ ದರಗಳನ್ನು ಸೇರಿಸಿ ನಮೂದಿಸಬೇಕು. ಇಲ್ಲದಿದ್ದಲ್ಲಿ ಹಾಲಿ ಸೂಚಿಸಿರುವ ದರವು ಎಲ್ಲವೂ ಸೇರಿರುವುದಾಗಿ ಭಾವಿಸಲಾಗುವುದು.
- 7) ಸಂಸ್ಥೆಯವರು ಪೂರೈಸುವ ಸಾಮಗ್ರಿಗಳು ಕಳಪೆ ಗುಣಮಟ್ಟದಾಗಿದ್ದಲ್ಲಿ ಅಥವಾ ಸಮರ್ಪಕವಾಗಿ ಸರಬರಾಜು ಮಾಡದ ಕಾರಣಗಳಿಂದ ಸರಬರಾಜು ಮಾಡುವ ಸಾಮಗ್ರಿಗಳನ್ನು ಸ್ವೀಕರಿಸುವ/ನಿರಾಕರಿಸುವ ಹಕ್ಕನ್ನು ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ, ದಕ್ಷಿಣ ಕನ್ನಡ ಇದರ ಖರೀದಿ ಸಮಿತಿ ಹೊಂದಿರುತ್ತದೆ.
- 8) ಇದು ಏಟಂವಾರು Manual ಟೆಂಡರ್ ಆಗಿರುತ್ತದೆ.
- 9) ಸಂಸ್ಥೆಗಳು ಸಮಾನ ದರವನ್ನು ನಮೂದಿಸಿದ್ದಲ್ಲಿ ಚೀಟಿ ಎತ್ತುವಿಕೆ ಪ್ರಕ್ರಿಯೆ ನಡೆಸಿ ಎಲ್-1 ದರ ಆಯ್ಕೆ ಮಾಡಲಾಗುವುದು.
- 10) ಸಾಮಗ್ರಿಗಳನ್ನು ಆದೇಶ ನೀಡಿದ 15 ದಿನಗಳ ಒಳಗಾಗಿ ಸುಳ್ಳುಬಂಟ್ವಾಳ, ಪುತ್ತೂರು ಮತ್ತು ಬೆಳ್ತಂಗಡಿ ತಾಲೂಕು ಆಸ್ಪತ್ರೆಯ Block Public Health Unit(BPHU)ಗಳಿಗೆ ಸರಬರಾಜು ಮಾಡತಕ್ಕದ್ದು.
- 11) ಅನುದಾನ ಲಭ್ಯವಿದ್ದಲ್ಲಿ 10 ದಿನಗಳ ಒಳಗಾಗಿ ಪಾವತಿ ಮಾಡಲಾಗುವುದು. ಅಗತ್ಯ ಅನುದಾನ ಲಭ್ಯವಿಲ್ಲದಿದ್ದಲ್ಲಿ ಅನುದಾನ ಸ್ವೀಕೃತವಾದ ನಂತರ ಬಿಲ್ ಪಾವತಿ ಮಾಡುವ ಬಗ್ಗೆ ಸಂಸ್ಥೆಯ ಒಪ್ಪಿಗೆ ಇರುತ್ತದೆ.
- 12) ಸಂಸ್ಥೆಯವರು ಮೇಲೆ ತಿಳಿಸಿದ ನಿಬಂಧನೆಗಳ ಹೊರತಾಗಿಯೂ ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ, ದಕ್ಷಿಣ ಕನ್ನಡ ಇದರ ಖರೀದಿ ಸಮಿತಿಯ ನಿರ್ಧಾರವೇ ಅಂತಿಮವಾಗಿರುತ್ತದೆ.

ಈ ಮೇಲಿನ ಎಲ್ಲಾ ಷರತ್ತುಗಳನ್ನು ಓದಿ ಅರ್ಥೈಸಿದ್ಧೇನೆ/ದ್ದೇವೆ ಮತ್ತು ಈ ಎಲ್ಲಾ ಷರತ್ತುಗಳಿಗೆ ಬದ್ಧನಾಗಿರುತ್ತೇನೆ/ತ್ತೇವೆ.

ಸಂಸ್ಥೆಯವರ ಸಹಿ ಮತ್ತು ಮೊಹರು

**Supply of Phlebotomy chair ,Three bucket mop trolley system & Bio-Medical
waste segregation trolley under BPHU Programme**

Item Specification & Quantity Details

Sl.No	Particulars	Sullia	Beltangady	Puttur	Bantwal	Total
1	Phlebotomy chair	1	1	-	-	2
2	Three bucket mop trolley system	1	1	1	1	4
3	Bio-Medical waste segregation trolley	1	1	1	1	4

ಸಹಿ/-
ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ
ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ, ಸೊಸೈಟಿ
ದ.ಕ ಮಂಗಳೂರು.

ಸಂಸ್ಥೆಯವರ ಸಹಿ ಮತ್ತು ಮೊಹರು

I. Technical specifications for Phlebotomy chair:

1. Material: MS powder coated
2. Should be designed for blood collection.
3. Should accommodate 1 person.
4. Should have a load bearing capacity of at least 150 kg.
5. Should have an adjustable back rest.
6. Should have a broad seat that can accommodate overweight or obese patients.
7. Should have broad arm rests on both sides of the chair with upholstery to comfortably rest arm for blood sample collection.
8. The arm rest should be adjustable in both vertical and horizontal plane.
9. Should have a non-slippery, comfortable, and adjustable foot rest. The chair shall have a provision to attach a sharp bin.
10. Should have head rest.
11. The chair should be covered with stain proof and washable/wipe clean material.
12. Should not have any sharp or pointed edges or corners.
13. Should have a firm and non-slippery legs/bottom to prevent displacement of the chair due to any jerky movements of the patient
14. One-year comprehensive WARRANTY should be provided for the quoted model.



II. Technical Specification of Three-Bucket Mop Trolley System :

1. Scope
Supply of three-bucket mop trolley system suitable for wet mopping and cleaning in hospitals, laboratories, and other healthcare facilities, with emphasis on hygiene control and ease of disinfection.
2. Construction
 - a. Heavy-duty trolley frame made of corrosion-resistant material
 - b. Smooth, non-porous surfaces for easy cleaning and disinfection
 - c. Stable design with no sharp edges

3. Buckets
 - a. Minimum three (03) detachable buckets
 - b. Capacity: Approx. 20–30 litres per bucket
 - c. Made of impact-resistant, chemical-resistant plastic
 - d. Different colors for segregation (e.g., clean water, dirty water, disinfectant)
 - e. Easy to remove, clean, and refill
4. Wringer
 - a. Heavy-duty wringer suitable for hospital-grade wet mops
 - b. Lever-operated wringing mechanism
 - c. Designed to reduce water splashing and operator effort
 - d. Securely mounted on trolley
5. Mobility
 - a. Four (04) swivel castor wheels, non-marking type
 - b. Minimum wheel diameter: 75 mm
 - c. At least two castors fitted with brakes
 - d. Smooth movement suitable for corridors and patient areas
6. Ergonomics & Safety
 - a. Ergonomic push handle for easy maneuvering
 - b. Stable under full load, resistant to tipping
 - c. Designed to minimize operator strain
7. Hygiene & Compatibility
 - a. Compatible with hospital disinfectants and cleaning chemicals
 - b. Materials shall not absorb moisture or odors
 - c. Suitable for infection control cleaning protocols
8. Warranty
 - a. Minimum one (01) year warranty against manufacturing defects
9. Acceptance Criteria

The supplied three-bucket mop trolley system shall be accepted if it meets the following:

1. Visual & Physical Inspection
 1. Frame is sturdy, corrosion-resistant, and free from sharp edges
 2. Buckets are intact, color-coded, and securely seated
 3. Wringer operates smoothly without excessive force
2. Functional Testing
 1. Castor wheels rotate freely and brakes function effectively
 2. Trolley moves smoothly when fully loaded with water
 3. Wringer effectively squeezes mop without water splashing
3. Hygiene Compliance
 1. Buckets and frame surfaces are easy to clean and disinfect
 2. Materials show no degradation when exposed to standard hospital disinfectants
4. Stability & Safety
 1. Trolley remains stable during movement and wringing operations
 2. No leakage from buckets during normal use
5. Documentation
 1. Manufacturer's product datasheet submitted, 2. Warranty certificate provided



III. Technical Specification of Biomedical Waste Segregation Trolley – 4 Bin (Hospital / Laboratory Use)

1. Scope

Supply of mobile biomedical waste segregation trolley with four (04) color-coded bins, suitable for collection and internal transportation of biomedical waste in hospitals, laboratories, and healthcare facilities, in compliance with Biomedical Waste

Management Rules.

2. Trolley Construction

- a. Frame made of heavy-duty, corrosion-resistant material (e.g., stainless steel or powder-coated metal)
- b. Rigid, stable structure suitable for continuous hospital use
- c. Smooth, non-porous surfaces for easy cleaning and disinfection
- d. No sharp edges or protrusions

3. Waste Bins

- a. Number of bins: Four (04)
- b. Individual bin capacity: Approximately 16 litres each
- c. Material: Durable, impact-resistant, non-toxic plastic
- d. Leak-proof and easy to clean
- e. Fitted with hinged lids

4. Color Coding & Marking

- a. Bins shall be color-coded and labelled as per Biomedical Waste Management

Rules:

- b. Yellow
- c. Red
- d. White (Translucent)
- e. Blue

- f. Each bin to display:
 - i. Appropriate biohazard symbol
 - ii. Waste category indication (printed or sticker)
- 5. Mobility
 - a. Four (04) swivel castor wheels for smooth movement
 - b. Minimum wheel diameter: 75 mm
 - c. Non-marking, floor-safe wheels
 - d. At least two wheels with locking/brake mechanism
- 6. Hygiene & Safety
 - a. Compatible with hospital disinfectants and cleaning agents
 - b. Design minimizes manual handling of waste
 - c. Suitable for infection control and segregation at source
- 7. Warranty
 - a. Minimum one (01) year warranty against manufacturing defects
- 8. Acceptance Criteria

The biomedical waste segregation trolley shall be accepted upon meeting the following:

 - a. Visual & Physical Inspection
 - i. Frame is sturdy, corrosion-resistant, and properly finished
 - ii. Bins are intact, leak-proof, and correctly color-coded
 - iii. Biohazard symbols and labels are clearly visible
 - b. Functional Testing
 - i. Wheels rotate freely and brakes function effectively
 - ii. Trolley moves smoothly when bins are fully loaded
 - iii. Bin lids open and close properly without breakage
 - c. Compliance Check
 - i. Color coding matches BMW Management Rules
 - ii. Materials suitable for healthcare waste handling
 - iii. Product free from manufacturing defects
 - d. Documentation
 - i. Product datasheet submitted
 - ii. User manual provided
 - iii. Warranty certificate provided



Model Bid form

Firm's letterhead

Date:

To,
The Member Secretary,
District Health and Family Welfare Society,
D.K, Mangaluru.
Sir,

Supply of Phlebotomy chair ,Three bucket mop trolley system & Bio-Medical waste segregation trolley under BPHU Programme

Ref No: DPMU/DK/BPHU /15th FC/94D/2025-26,Dated 11/02/2026

Financial Bid

Sl.No	Item Description	Rate	Quantity	Total
1	Phlebotomy chair		2	
2	Three bucket mop trolley system		4	
3	Bio-Medical waste segregation trolley		4	

Detailed Technical Specification is attached along with this tender document
Rate should be Inclusive of GST,TDS, Transportation cost.

Bidder Seal & sign
