



ಕರ್ನಾಟಕ ಸರ್ಕಾರ

ರಾಷ್ಟ್ರೀಯ ಆರೋಗ್ಯ ಅಭಿಯಾನ

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ (ರಿ)

ಜಿಲ್ಲಾ ಯೋಜನಾ ನಿರ್ವಹಣಾ ಘಟಕ, ಪಿಡಬ್ಲ್ಯುಡಿ ಬಿಲ್ಡಿಂಗ್, ನೆಹರೂ ಮೈದಾನ್ ರಸ್ತೆ ದ.ಕ. ಮಂಗಳೂರು

ಇಮೇಲ್: rchodkannada@gmail.com

ಫೋನ್ ನಂಬ್ರ: 0824-2424501

ನಂ: DPMU/DK/15thFC/NC/86/2024-25/Volume-02

ದಿನಾಂಕ: 16/04/2026

15ನೇ ಹಣಕಾಸು ಆಯೋಗದಡಿಯಲ್ಲಿ

‘ನಮ್ಮ ಕ್ಲಿನಿಕ್’ ಮತ್ತು ತಾಲೂಕು

ಆಸ್ಪತ್ರೆಗಳಿಗೆ ಅಗತ್ಯವಿರುವ

ಸಾಧನೋಪಕರಣಗಳನ್ನು

ಪೂರೈಸಲು ಅಲ್ಪಾವಧಿ Manual

ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಆಹ್ವಾನ



ಕರ್ನಾಟಕ ಸರ್ಕಾರ

ರಾಷ್ಟ್ರೀಯ ಆರೋಗ್ಯ ಅಭಿಯಾನ

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೊಸೈಟಿ (೦)

ಜಿಲ್ಲಾ ಯೋಜನಾ ನಿರ್ವಹಣಾ ಘಟಕ, ಪಿಡಬ್ಲ್ಯುಡಿ ಬಿಲ್ಡಿಂಗ್, ನೆಹರೂ ಮೈದಾನ ರಸ್ತೆ ದ.ಕ. ಮಂಗಳೂರು

ಇಮೇಲ್: rchodkannada@gmail.com

ಫೋನ್ ನಂಬ್ರ: 0824-2424501

ನಂ: DPMU/DK/15thFC/NC/86/2024-25/Volume-02

ದಿನಾಂಕ: 16/04/2026

Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಪ್ರಕಟಣೆ

15ನೇ ಹಣಕಾಸು ಆಯೋಗದಡಿಯಲ್ಲಿ ನಗರ ಸ್ಥಳೀಯ ಸಂಸ್ಥೆಗಳಿಗೆ ಅನುಮೋದನೆಗೊಂಡ ಅನುದಾನದಡಿಯಲ್ಲಿ ದಕ್ಷಿಣ ಕನ್ನಡ ಜಿಲ್ಲೆಯಲ್ಲಿ ಈಗಾಗಲೇ ಕಾರ್ಯನಿರ್ವಹಿಸುತ್ತಿರುವ ನಮ್ಮ ಕ್ಲಿನಿಕ್ ಆರೋಗ್ಯ ಕೇಂದ್ರಗಳಿಗೆ ಮತ್ತು ತಾಲೂಕು ಆಸ್ಪತ್ರೆಗಳಿಗೆ ಅಗತ್ಯವಿರುವ ಸಾಧನೋಪಕರಣಗಳನ್ನು ಪೂರೈಸಲು ಅರ್ಹ ಸಂಸ್ಥೆಗಳಿಂದ Manual ಟೆಂಡರ್/ ದರಪಟ್ಟಿ ಆಹ್ವಾನಿಸಲಾಗಿದೆ. ಅರ್ಹ ಸಂಸ್ಥೆಗಳು ಉಚಿತ ಟೆಂಡರ್ ನಮೂನೆಗಳಿಗಾಗಿ ಈ ಕಛೇರಿಯ ವೆಬ್‌ಸೈಟ್ <https://www.dkhfw.in> ತಂತ್ರಾಂಶದಲ್ಲಿ ನಮೂನೆಯನ್ನು ಪಡೆದು ಭರ್ತಿಗೊಳಿಸಿದ Manual ಟೆಂಡರ್/ ದರಪಟ್ಟಿ ನಮೂನೆಯನ್ನು Technical Document ನಲ್ಲಿ ತಿಳಿಸಲಾದ ಎಲ್ಲಾ ದಾಖಲೆಗಳ ಪ್ರತಿಯೊಂದಿಗೆ The Member Secretary, District Health and Family Welfare Society, DK, Mangaluru(ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ, ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ.ಸೊಸೈಟಿ, ದ.ಕ.ಮಂಗಳೂರು)ಇವರ ಹೆಸರಿನಲ್ಲಿ ಮೊಹರಾದ ಲಕೋಟಿಯಲ್ಲಿಟ್ಟು ದಿನಾಂಕ 23/04/2026 ರ ಪೂರ್ವಾಹ್ನ 11.00 ಗಂಟೆಯೊಳಗಾಗಿ ಈ ಕಛೇರಿಗೆ ಸಲ್ಲಿಸತಕ್ಕದ್ದು. Manual ಟೆಂಡರ್/ ದರಪಟ್ಟಿ ನಮೂನೆಯನ್ನು ದಿನಾಂಕ 23/04/2026 ರ ಪೂರ್ವಾಹ್ನ 11.30 ಗಂಟೆಗೆ ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ.ಕಛೇರಿಯಲ್ಲಿ ತೆರೆಯಲಾಗುವುದು.

ಸಹಿ/-

ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ,

ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ ಸೊಸೈಟಿ,

ದ.ಕ.ಮಂಗಳೂರು

ಬಿಡ್‌ದಾರರ ಸಹಿ ಮತ್ತು ಮೊಹರು

**Supply of Equipments under 15th Finance ULB Budget in Dakshina
Kannada District**

Technical Document

Sl.No	Particulars	Details to be Filled by Tenderer (Attached/Not attached)
1	Name of the Firm, Firm full address Office Phone, Fax, and Mobile No and E-Mail ID (in Firm's Letter Head)	
2	Attested Copy of firm Registration Certificate	
3	Attested Copy of GST Registration Certificate	
4	Attested Copy of PAN Card	
5	Technical literature of the product quoted (Brochures)	

ಟೆಂಡರ್‌ದಾರರ ಸಹಿ ಮತ್ತು ಮೊಹರು

15ನೇ ಹಣಕಾಸು ಆಯೋಗದಡಿಯಲ್ಲಿ ನಮ್ಮ ಕ್ಲಿನಿಕ್ ಮತ್ತು ತಾಲೂಕು ಆಸ್ಪತ್ರೆಗಳಿಗೆ ಸಾಧನೋಪಕರಣಗಳನ್ನು ಪೂರೈಸಲು ನಿಬಂಧನೆಗಳು

- 1) ಅಂದಾಜು ವೆಚ್ಚ ಮೊತ್ತ 3,20,000/- ಆಗಿರುತ್ತದೆ.
- 2) Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಪರಿಶೀಲನೆಯ ಬಗ್ಗೆ ಮತ್ತು ಈ Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಬಗ್ಗೆ ಇತರೇ ಯಾವುದೇ ದೂರುಗಳಿದ್ದಲ್ಲಿ ಸದರೀ ಪ್ರಕ್ರಿಯೆ ನಡೆದ 1 ವಾರದೊಳಗೆ ಮಾತ್ರವೇ ಮೇಲ್ಮನವಿ ಪ್ರಾಧಿಕಾರಕ್ಕೆ ಮನವಿ ಸಲ್ಲಿಸಬೇಕು.ತದನಂತರ ಬಂದ ಯಾವುದೇ ಮನವಿಯನ್ನು ಪರಿಗಣಿಸಲಾಗುವುದಿಲ್ಲ.
- 3) ಯಶಸ್ವಿ ಸಂಸ್ಥೆಯು/ಇಲಾಖೆಯು ಒಪ್ಪಂದದಿಂದ ಹಿಂದೆ ಸರಿಯಾದಾದರೆ 7 ದಿನಗಳ ಮುಂಚಿತವಾಗಿ ನೋಟೀಸ್‌ನ್ನು ಬರವಣಿಗೆಯಲ್ಲಿ ನಿಡತಕ್ಕದ್ದು. ಇಲ್ಲದಿದ್ದಲ್ಲಿ ಅಗತ್ಯ ಕ್ರಮ ಕೈಗೊಳ್ಳಲಾಗುವುದು.
- 4) ಸಂಸ್ಥೆಯವರು ಸೂಕ್ತವಾಗಿ Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಸಲ್ಲಿಕೆ ಮಾಡಬೇಕು.ಯಾವುದೇ ತಪ್ಪುಗಳು ಸಂಭವಿಸಿದಲ್ಲಿ ಇಲಾಖೆಯು ಜವಾಬ್ದಾರಿ ಆಗಿರುವುದಿಲ್ಲ.
- 5) ಯಾವುದೇ ಸೂಚನೆಗಳನ್ನು ನೀಡದೇ Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿ ಅನ್ನು ರದ್ದುಪಡಿಸಲು/ಅಂಗೀಕರಿಸಲು/ತಿರಸ್ಕರಿಸುವ ಹಕ್ಕು ಇಲಾಖೆಗೆ ಇರುತ್ತದೆ.
- 6) Manual ಟೆಂಡರ್/ದರಪಟ್ಟಿಯಲ್ಲಿ ನಮೂದಿಸುವ ದರವನ್ನು ಮಾತ್ರ ಪರಿಗಣಿಸಲಾಗುವುದು.ಆದುದರಿಂದ ಸಂಸ್ಥೆಯವರು ದರಗಳನ್ನು ಸೂಚಿಸುವಾಗ ಸಾಗಾಟದ ವೆಚ್ಚ, ತೆರಿಗೆ ದರಗಳನ್ನು ಸೇರಿಸಿ ಸಮೂದಿಸಬೇಕು. ಇಲ್ಲದಿದ್ದಲ್ಲಿ ಹಾಲಿ ಸೂಚಿಸಿರುವ ದರವು ಎಲ್ಲವೂ ಸೇರಿರುವುದಾಗಿ ಭಾವಿಸಲಾಗುವುದು.
- 7) ಅನುದಾನಗಳ ಲಭ್ಯತೆಗನುಗುಣವಾಗಿ ಸಾಧನೋಪಕರಣಗಳನ್ನು ಖರೀದಿಸಲಾಗುವುದು ಎಂಬ ನಿಬಂಧನೆಗೊಳಪಟ್ಟಿರುತ್ತದೆ.
- 8) ಸಂಸ್ಥೆಯವರು ಪೂರೈಸುವ ಸಾಧನೋಪಕರಣಗಳು ಕಳಪೆ ಗುಣಮಟ್ಟದ್ದಾಗಿದ್ದಲ್ಲಿ ಅಥವಾ ಸಮರ್ಪಕವಾಗಿ ಸರಬರಾಜು ಮಾಡದ ಕಾರಣಗಳಿಂದ ಸರಬರಾಜು ಮಾಡುವ ಸಾಧನೋಪಕರಣಗಳನ್ನು ಸ್ವೀಕರಿಸುವ/ನಿರಾಕರಿಸುವ ಹಕ್ಕನ್ನು ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ.ಸೊಸೈಟಿ,ದ.ಕ ಇದರ ಖರೀದಿ ಸಮಿತಿ ಹೊಂದಿರುತ್ತದೆ.
- 9) ಸಾಧನೋಪಕರಣಗಳನ್ನು ಆದೇಶ ನೀಡಿದ 10 ದಿನಗಳ ಒಳಗಾಗಿ ಆದೇಶದೊಂದಿಗೆ ನೀಡಿದ ವಿಳಾಸಕ್ಕೆ ಸರಬರಾಜು ಮಾಡತಕ್ಕದ್ದು.
- 10) ಅನುದಾನ ಲಭ್ಯವಿದ್ದಲ್ಲಿ 10 ದಿನಗಳ ಒಳಗಾಗಿ ಪಾವತಿ ಮಾಡಲಾಗುವುದು.ಅಗತ್ಯ ಅನುದಾನ ಲಭ್ಯವಿಲ್ಲದಿದ್ದಲ್ಲಿ ಅನುದಾನ ಸ್ವೀಕೃತವಾದ ನಂತರ ಬಿಲ್ ಪಾವತಿ ಮಾಡುವ ಬಗ್ಗೆ ಸಂಸ್ಥೆಯ ಒಪ್ಪಿಗೆ ಇರುತ್ತದೆ.
- 11) ಕೋರಿರುವ ಸಾಧನೋಪಕರಣಗಳು ಎಷ್ಟೇ ಪ್ರಮಾಣದಲ್ಲಿ ಕೇಳಿದರೂ 7 ದಿನಗಳ ಒಳಗೆ ಸರಬರಾಜು ಮಾಡಲು ಬದ್ಧರಾಗಿರತಕ್ಕದ್ದು.
- 12) ಸಂಸ್ಥೆಯವರು ಮೇಲೆ ತಿಳಿಸಿದ ನಿಬಂಧನೆಗಳ ಹೊರತಾಗಿಯೂ ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ.ಸೊಸೈಟಿ,ದ.ಕ ಇದರ ಖರೀದಿ ಸಮಿತಿಯ ನಿರ್ಧಾರವೇ ಅಂತಿಮವಾಗಿರುತ್ತದೆ.

ಈ ಮೇಲಿನ ಎಲ್ಲಾ ಶರತ್ತುಗಳನ್ನು ಓದಿ ಅರ್ಥೈಸಿದ್ದೇನೆ/ದ್ದೇವೆ ಮತ್ತು ಈ ಎಲ್ಲಾ ಷರತ್ತುಗಳಿಗೆ ಬದ್ಧನಾಗಿರುತ್ತೇನೆ/ತ್ತೇವೆ.

ಸಂಸ್ಥೆಯವರ ಸಹಿ ಮತ್ತು ಮೊಹರು

**Supply of Equipments under 15th Finance ULB Fund in Dakshina
Kannada District**

Item Specification & Quantity Details

Sl.No	Item Name	Specification	Qty
1	CENTRIFUGE	<p>Used for Biochemical & Analytical labs for Haematocrit, blood corpuscle presentage ,serum analysis,precipitate separation & blood group matching. 1).Speed:Maximum range 4000 to 6000 RPM</p> <p>2). Reciprocating Centerfugal force (RCF):3000 to 3500</p> <p>3) Minimum Capacity 240ml</p> <p>4)Digital Timer Range 0 to 59 Minutes</p> <p>5)Auto Lid Interlock to Prevent opening while running centrifuge with emergency lidlock release.</p> <p>6)Motor inbalance detector future-desirable</p> <p>7)Microprocessor with digital display</p> <p>8)Dynamic break for quick declarations</p> <p>9)Stainless Steel Chamber easy to clean</p> <p>10) Hinges to prevent door falling</p> <p>11)Rotor size 16 X 15ml</p> <p>12)Rotor should be auto cleavable</p> <p>13) Brushless rotor Capacity 120 ml or above</p> <p>14) Power Requirement - 220-240 V/50Hz</p> <p>15)Power consumption 400 to 500 Watts</p> <p>16) Accessories- Rubber adapter should be provider for the use of vacutainer for 3ml and 5ml</p> <p>17)Operating Condition- Capable of operating continusly in ambient temperature of 10 to 50 deg C and Relative humidity of 15 to 90% in ideal circumstances</p> <p>18) Storage Condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and Relative humidity of 15 to 90%</p> <p>19)Should be FDA/CE/BIS approved product</p> <p>20) Manufacturer and supplier should have ISO 13485 certification for quality standards</p> <p>21)Electrical safty conforms to the standards for electrical safty IEC 60601- General requirements(or equivalent BIS standard</p> <p>22)Shall meet internationally recognasied standards for electromagnetic compatibility (EMC)for electro medical equipment:61326-1</p> <p>23) Certified to be compliant with IEC 61010-1,IEC 61010-2-40 for safety.</p> <p>24)availablity of 5 amp socket</p> <p>25) Certification of calibration and inspection from the manufacturer</p> <p>26)3 year Warranty</p> <p>27)CMC 4 Year 2 PM and Calibration Visits Annually all Breakdown call to be attended within 24 hr of registration</p>	12

2	PIPETTE SET	<p>1.Micropipettes Single Channel of Variable Volume (0.1-2ul, 1-10 ul,2-20 ul,20-200 ul and 100-1000 ul)</p> <p>2.Extremely light pipetting action by tension free apring* mechanism</p> <p>3. Fully autoclavable (121 C):UV resistant material</p> <p>4. Increments at least 0.1 ul fine adjustment for pipettes expect for 100-1000 ul volume pipette where increment can be read as 0.5 ul</p> <p>5. Suitable for all brands of tips</p> <p>6. Adjustable for variable volume fixation of adjustable volumes</p>	12
3	RETINOSCOPE	<p>* External Focusing Sleeve - Unique planetary gear system which allows for easy adjustment no matter what size hand or how an instrument is held. Continuous 360 rotation. Should maintain the same plane of focuse during rotation</p> <p>* Dust- Free optics - Housings and glass cover one the front keep the instrument cleaner longer(multi coated optics)</p> <p>* Crossed Linear polarizing filter which reduces glare from lenses allowing retinoscopy to be performed closer to theaxis of the correcting lenses</p> <p>*Fixation Cards- Must provide Card that easily attach to increase the case with which dynamic retinoscopy is performed</p> <p>* Optics- Glare and shadows should be eliminated for a clearer and more precise view</p> <p>* Interchange ability * By simple changing the lamp,the streak retinoscope should be converted ta a spot retinoscope</p> <p>* Should be interchangeable to plane mirror and concave mirror mode by sleeve movement</p> <p>*Should allow one-hand operation for streak focus</p> <p>*Should be ISO/FDA/CE/CDSO or BIS Approved</p>	2

ಸದಸ್ಯ ಕಾರ್ಯದರ್ಶಿ,
ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕು.ಕ ಸೊಸೈಟಿ,
ದ.ಕ,ಮಂಗಳೂರು

ಸಂಸ್ಥೆಯವರ ಸಹಿ ಮತ್ತು ಮೊಹರು

Model Bid Form
Firm's Letterhead

Date:

To,
The Member Secretary,
District Health and Family Welfare Society,
D.K, Mangaluru

Sir,

Supply of Equipments under 15th Finance ULB Fund.
Financial Bid

Sl.No	Item Name	Specification	Qty	Rate	GST	Total Amount
1	CENTRIFUGE	Used for Biochemical & Analytical labs for Haematocrit, blood corpuscle presentage ,serum analysis, precipitate separation & blood group matching. 1).Speed:Maximum range 4000 to 6000 RPM 2). Reciprocating Centerfugal force (RCF):3000 to 3500 3) Minimum Capacity 240ml 4)Digital Timer Range 0 to 59 Minutes 5)Auto Lid Interlock to Prevent opening while running centrifuge with emergency lidlock release. 6)Motor inbalance detector future-desirable 7)Microprocessor with digital display 8)Dynamic break for quick declarations 9)Stainless Steel Chamber easy to clean 10) Hinges to prevent door falling 11)Rotor size 16 X 15ml 12)Rotor should be auto cleavable 13) Brushless rotor Capacity 120 ml or above 14) Power Requirement - 220-240 V/50Hz 15)Power consumption 400 to 500 Watts 16) Accessories- Rubber adapter should be provider for the use of vacutainer for 3ml and 5ml 17)Operating Condition- Capable of operating continusly in ambient temperature of 10 to 50 deg C and Relative humidity of 15 to 90% in ideal circumstances 18) Storage Condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and Relative humidity of 15 to 90% 19)Should be FDA/CE/BIS approved product 20) Manufacturer and supplier should have ISO 13485 certification for quality standards 21)Electrical safty conforms to the standards for electrical safty IEC 60601- General requirements(or equivalent BIS standard	12			

		<p>22) Shall meet internationally recognised standards for electromagnetic compatibility (EMC) for electro medical equipment: 61326-1</p> <p>23) Certified to be compliant with IEC 61010-1, IEC 61010-2-40 for safety.</p> <p>24) availability of 5 amp socket</p> <p>25) Certification of calibration and inspection from the manufacturer</p> <p>26) 3 year Warranty</p> <p>27) CMC 4 Year 2 PM and Calibration Visits Annually all Breakdown call to be attended within 24 hr of registration</p>				
2	PIPETTE SET	<p>1. Micropipettes Single Channel of Variable Volume (0.1-2ul, 1-10 ul, 2-20 ul, 20-200 ul and 100-1000 ul)</p> <p>2. Extremely light pipetting action by tension free apring mechanism</p> <p>3. Fully autoclavable (121 C): UV resistant material</p> <p>4. Increments at least 0.1 ul fine adjustment for pipettes expect for 100-1000 ul volume pipette where increment can be read as 0.5 ul</p> <p>5. Suitable for all brands of tips</p> <p>6. Adjustable for variable volume fixation of adjustable volumes</p>	12			
3	RETINOSCOPE	<p>* External Focusing Sleeve - Unique planetary gear system which allows for easy adjustment no matter what size hand or how an instrument is held. Continuous 360 rotation. Should maintain the same plane of foecuse during rotation</p> <p>* Dust- Free optics - Housings and glass cover one the front keep the instrument cleaner longer (multi coated optics)</p> <p>* Crossed Linear polarizing filter which reduces glare from lenses allowing retinoscopy to be performed closer to the axis of the correcting lenses</p> <p>* Fixation Cards- Must provide Card that easily attach to increase the ease with which dynamic retinoscopy is performed</p> <p>* Optics- Glare and shadows should be eliminated for a clearer and more precise view</p> <p>* Interchange ability * By simple changing the lamp, the streak retinoscope should be converted ta a spot retinoscope</p> <p>* Should be interchangeable to plane mirror and concave mirror mode by sleeve movement</p> <p>* Should allow one-hand operation for streak focus</p> <p>* Should be ISO/FDA/CE/CDSO or BIS Approved</p>	2			